

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Hightstown Police Department 415A Mercer Street Hightstown, NJ 08520</b>	2. Contact: <b>Lt. Frank Gendron</b>  3. Phone Number: <b>609-448-1234 x507</b>	4. Date: <b>05/10/2013</b>  5. County: <b>Mercer</b>
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>N/A</b></p>	7. Simulator Component Serial Number: <p style="text-align: center;"><b>DDUL S3-0183</b></p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>

9. Reason for Service:  
**Complaint received that the simulator was inoperable.**

10. Comments:  
**Operational condition of the simulator was assessed. I observed the heater to be discolored and the baffle to do be melted.**  
  
**Simulator placed out of service and returned to Draeger Safety Diagnostics for evaluation.**

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Tpr. II Michael P. Gibson #6353**  
 Name & Badge Number (Print)

*Michael P. Gibson #6353 5/10/13*  
 Signature Date

# Packing Slip

Customer no    Order No    Order date  
 150054156    10625936    05/17/2013

Packing slip no.    Ship date  
 80775003    05/17/2013  
 Please reference on inquiries

ship to  
 150054156  
 BOROUGH OF HIGHTSTOWN POLICE DEPT.  
 ATTN: SGT. MARK NIRO  
 MERCER COUNTY  
 415A MERCER STREET  
 HIGHTSTOWN, NJ 08520  
 USA

bill to  
 [Empty box for bill to information]

Your Purchase Order Number  
 05/17/2013  
 RECERT/RECAL  
 Any questions? Please contact:  
 MARA HERRERA

Customer  
 BOROUGH OF HIGHTSTOWN POLICE DEPT.  
 ATTN: SGT. MARK NIRO  
 MERCER COUNTY  
 415A MERCER STREET  
 HIGHTSTOWN, NJ 08520  
 USA

Your sales person  
 Don Pouliot  
 Phone: 973-398-3228

Delivery terms  
 FA  
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Ground			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		SIM: DDULS3-0183 PROBE: DDWJP2-055			
0010	1 EA	MPCERT CERTIFICATION CHARGE SIMULATOR Ordered / Back ordered : 1 / 0 EA			
0020	1 EA	MPCALTP CAL CHARGE, SIM PROBE Ordered / Back ordered : 1 / 0 EA			

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150054156    10625936    05/17/2013

Packing slip no.    Ship date  
80775003    05/17/2013

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

# Packing Slip

Customer no    Order No    Order date  
150054156    10622719    05/03/2013

Packing slip no.    Ship date  
80771381    05/03/2013

Please reference on inquiries

ship to

150054156  
BOROUGH OF HIGHTSTOWN POLICE DEPT.  
ATTN: SGT. MARK NIRO  
MERCER COUNTY  
415A MERCER STREET  
HIGHTSTOWN, NJ 08520  
USA

bill to

Your Purchase Order Number

05/03/2013  
RECERT/RECAL  
Any questions? Please contact:  
AMANDA DILLEY

Customer

BOROUGH OF HIGHTSTOWN POLICE DEPT.  
ATTN: SGT. MARK NIRO  
MERCER COUNTY  
415A MERCER STREET  
HIGHTSTOWN, NJ 08520  
USA

Your sales person

Don Pouliot  
Phone: 973-398-3228

Delivery terms

FA  
FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Ground			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		SIM: DDULS3-0182 ; PROBE: DDUJP2-035 REC'D ON 5-2-13			
0010	1 EA	MPCERT CERTIFICATION CHARGE SIMULATOR Ordered / Back ordered : 1 / 0 EA			
0020	1 EA	MPCALTP CAL CHARGE, SIM PROBE Ordered / Back ordered : 1 / 0 EA			

**Draeger Safety Diagnostics Inc.**  
Accounting Address:  
101 Technology Drive  
Pittsburgh, PA 15275  
Tel: (412) 788-5537  
Fax: (412) 788-5598

**Remit Wire Transfers To:**  
Citizens Bank  
Acct. # 6209426615  
Acct. Name: Draeger Safety Diagnostics Inc.  
Transit Number: 036076150  
Federal ID Number: 84-1600159

**REMIT TO:**  
Draeger Safety Diagnostics Inc.  
P.O. Box 200337  
Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150054156    10622719    05/03/2013

Packing slip no.    Ship date  
80771381    05/03/2013

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 1 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

# Dräger

## ALCOTEST® 7110 TEMPERATURE PROBE

### CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDUSP2-035

Certification date:

5-3-13

Next Certification due:


5-3-14

Probe Value

98

Draeger Safety Diagnostics, Inc.  
Technical Service Department

BC



# Dräger

### CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: \_\_\_\_\_

Serial Number:

DDUS3-082

Certification Date

5-3-13

Technician

BC

Re-Certification Due Date

5-3-14

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 00901 Calib. Date: 03/14/2013 Calib. No.: 00021  
Certification File No.: 00902 Cert. Date: 03/14/2013 Cert. No.: 00016  
Linearity File No.: 00903 Lin. Date: 03/14/2013 Lin. No.: 00016  
Solution File No.: 00913 Soln. Date: 05/10/2013 Soln. No.: 00154  
Sequential File No.: 00913 File Date: 05/10/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182  
Control Solution %: 0.100% Expires: 09/13/2014  
Solution Control Lot: 121107 Bottle No.: 0861

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:40D		
Control 1 EC	0.100%	12:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41D		
Control 2 EC	0.100%	12:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		
Control 3 EC	0.100%	12:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJP2 - 035 (MPA)

**Changed By:**

Last Name: GIBSON

First Name: MICHAEL

MI: P

Signature: Michael P. Gibson #6353

Badge No.: 6353

Date: 05/10/2013