

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
 Location: HIGHTSTOWN POLICE  
 Calibration File No.: 00757 Calib. Date: 09/16/2011 Calib. No.: 00017  
 Certification File No.: 00758 Cert. Date: 09/16/2011 Cert. No.: 00013  
 Linearity File No.: 00759 Lin. Date: 09/16/2011 Lin. No.: 00013  
 Solution File No.: 00760 Soln. Date: 09/16/2011 Soln. No.: 00120  
 Sequential File No.: 00760 File Date: 09/16/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183  
 Control Solution %: 0.100% Expires: 07/07/2013  
 Solution Control Lot: 11G093 Bottle No.: 0887

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:51D		
Control 1 EC	0.101%	14:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		
Control 2 EC	0.100%	14:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54D		
Control 3 EC	0.100%	14:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ P2-0559 *Handwritten signature*

**Changed By:**

Last Name: WEYMAN

First Name: HAROLD

MI: -

Signature: *Det. H. Weyman #6289*

Badge No.: 6289

Date: 09/16/2011